



**nexttec**<sup>TM</sup>  
DNA isolation systems

**Special Protocol**  
**DNA Isolation**  
*Fish*

by nexttec<sup>TM</sup> 1<sup>-Step</sup>

- nexttec<sup>TM</sup> cleanPlate96 -

**Cat. No. 10N.901**

**Cat. No. 10N.902**

**Cat. No. 10N.904**

**Cat. No. 10N.924**

Version 1.0

**For research only**

## Principle

nexttec™ 1-Step is the easiest handling and fastest DNA purification system containing a single buffer system and a one-step DNA purification after lysis.

Proteins, detergents and low molecular weight compounds are retained by the nexttec™ sorbent. DNA passes through the nexttec™ cleanPlate96 during a short, one-step purification procedure.

The obtained DNA is suitable for all common enzymatic reactions (restriction digests, real-time PCR, PCR, genotyping etc.).

## Kit contents

The kit contains all necessary reagents for lysis and subsequent DNA purification.

Component	Art.No. 10N.901	Art.No. 10N.902	Art.No. 10N.904	Art.No. 10N.924
Buffer G	15 ml	42 ml	65 ml	400 ml
Proteinase K	1.5 ml	3 ml	4.5 ml	27 ml
Prep Solution	40 ml	100 ml	150 ml	2x 450 ml
DTT (1,4- Dithio-DL-threitol)	0.5 ml	0.5 ml	1 ml	5 ml
nexttec™ cleanPlates96	1	2	4	24
nexttec™ deep-well plates	3	6	12	72
Sealing tapes	3	6	12	72
Alu sealing tapes	2	4	8	48

## nexttec™ service

For extending the application range to samples, which are difficult to lyse by the standard procedure, it is recommended to include optional components in the lysis buffer, to optimize the lysis time or to use different lysis volumes (e.g. robotic applications).

For detailed information, please get in contact with [service@nexttec.biz](mailto:service@nexttec.biz).

## Storage Conditions

During shipment all kit components are stable at room temperature. After arrival, **Proteinase K**, **Buffer G** and **Prep Solution** must be stored at **+2°C to +8°C**. After first opening freeze **DTT** at **-18°C to -25°C**.

**nexttec™ cleanPlates96** can be stored at **room temperature (+20°C to +25°C)**. If properly stored, see expiration date for the stability of the kit.

**Safety Information**

<b>Proteinase K</b>	Xn	R 36/37/38 R42/43;	S 23-24-26 36/37
<b>DTT</b>	Xn	R 22-36/37/38;	S 26-36

**Risk Phrases**

<b>R 22</b>	Harmful if swallowed
<b>R 36/37/38:</b>	Irritating to eyes, respiratory system and skin.
<b>R 42/43:</b>	May cause sensitisation by inhalation and skin contact.

**Safety Phrases**

<b>S 23:</b>	Do not breathe Gas/fumes/vapour/spray.
<b>S 24:</b>	Avoid contact with skin.
<b>S 26:</b>	In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.
<b>S 36/37:</b>	Wear suitable protective clothing and gloves.

When working with chemicals, always wear a suitable lab coat, disposable gloves and protective goggles. For more information please consult the appropriate material safety data sheets (MSDS).

**Before starting**

- **Equilibrate nexttec™ cleanPlates96**

E1	Add <b>350 µl Prep Solution</b> onto each well of a <b>nexttec™ cleanPlate96</b> . Incubate for at least <b>5 min</b> at room temperature and centrifuge at <b>350x g</b> for <b>1 min</b> or apply vacuum for <b>30 to 60 sec</b> to remove excess buffer.
E2	Discard the deep-well plate. Place the <b>nexttec™ cleanPlate96</b> onto a new deep-well plate. Use equilibrated <b>nexttec™ cleanPlates96</b> or store closed at <b>+2°C to +8°C</b> and use within one week.

- **Preheat an incubator to 56°C**

## **Protocol**

### **Lysis**

L1	Transfer <b>fish tissues (8 – 20 mg fresh weight)</b> to a deep-well plate.
L2	Add <b>140 µl Buffer G, 10 µl Proteinase K and 1.5 µl DTT*</b> to each sample. Incubate sample with shaking ( <b>56°C, 200 rpm, 20 min</b> ).

\*For Pre-Mixes see Technical Section.

### **Purification of DNA**

P	Transfer <b>100 µl</b> of the lysates to an <b>equilibrated nexttec™ cleanPlate96</b> . Incubate for <b>3 min</b> at <b>room temperature</b> . Centrifuge at <b>700x g</b> for <b>1 min</b> or apply vacuum for <b>1 min</b> .
	<b>The eluate contains the purified DNA!!</b>

### **Notes:**

---

---

---

---

---

---

---

---

---

---

---

---

---

---

---

---

---

---

## Technical Section

- **Preparation of Lysis Pre-Mixes**

	<b>Lysis Buffer LG:</b>	<b>1 sample</b>	<b>1 plate</b>	<b>2 plates</b>	<b>3 plates</b>	<b>4 plates</b>
<b>LG</b>	<b>Buffer G</b>	140 µl	15.4 ml	30.8 ml	46.2 ml	61.6 ml
	<b>Proteinase K</b>	10 µl	1.1 ml	2.2ml	3.3 ml	4.4 ml
	<b>DTT (optional)</b>	1.5 µl	165 µl	330 µl	495 µl	660 µl
	Mix by vortexing. Add <b>150 µl</b> of <b>Buffer LG</b> to each sample (L 2). The <b>Lysis Buffer LG</b> is stable for <b>1</b> working day, if stored at <b>+2°C to +8°C</b> .					

- **Determination of DNA concentration in nexttec™ 1-Step DNA preparations**

We recommend to determine the DNA concentration:

- Using the fluorescent dye Picogreen® or similar
- Comparing the fluorescence intensity of DNA bands of unknown concentration with standards, e.g. in ethidium bromide stained agarose gels.

Please notice:

The use of absorption measurement at 260 nm ( $A_{260}$ ) in a spectrophotometer (e.g. NanoDrop®) for determination of DNA concentration is system related not recommended. For details and possible workarounds for your specific application please contact:

[service@nexttec.biz](mailto:service@nexttec.biz).

- **Centrifugation**

For centrifugation at 350x g or 700x g use the settings for relative centrifugal force (RCF) of your centrifuge. Alternatively measure the distance of the nexttec™ cleanPlate96 to the centre of your rotor and calculate the necessary rotations per minute (e.g. rpm =  $299.07 \times \sqrt{350/r}$  ; r=radius in cm).

- **Vacuum Application**

Assemble the vacuum manifold as described in the manual. The flow rate should be between 15 L/min and 150 L/min.

**A regulation of vacuum is not necessary.**

## **Product Use Restriction**

nexttec™ 1-Step DNA Isolation Kit components were developed, designed and sold **for research purposes only**. They are suitable for in vitro uses only. No claim or representation is intended for use to identify any specific organism or of clinical use.

It is the responsibility of the user to verify the use of the nexttec™ 1-Step DNA Isolation Kit for a specific application as the performance characteristic of this kit has not been verified to a specific organism.

## **Troubleshooting, FAQ and Special Applications**

Product claims are subject to change. Therefore, please visit our website or contact our technical service team for troubleshooting guide, up-to-date protocols and latest applications on nexttec™ 1-Step products.

## **Contact Information**

Web: [www.nexttec.biz](http://www.nexttec.biz)

Email: [service@nexttec.biz](mailto:service@nexttec.biz)

Tel: +49 (0)214 / 869 15 15

Fax: +49 (0)214 / 869 15 28

Skype: nexttec1

## **Ordering Information**

For ordering information please visit our website [www.nexttec.biz](http://www.nexttec.biz) .

## **Distributor**

